

Task number

#81640

Testing ordered

30 MAY 26

Sample Received

10 JUL 26

Client

Peplab

Sample

SS-31

Manufacturer

Peplab

Batch

D3319047

Sample description

See picture or picture attached.

Tests requested

Assessment of a peptide vial or vials.

Results

SS-31

10.56 mg

Purity

99.507%

Comments

Analysis conducted

20 JUL 26

Verify ths test at www.eimi-institute.org/verify/
with the following unique key

L7X2M9Q4KPZA


Results





D3319047

Chromatogram



Purity: 99.507%

Method

Time	H2O + 0,1% TFA	ACN + 0,1% TFA
0 min	95%	5%
2 min	95%	5%
12 min	6%	94%
15 min	6%	94%